



ZPure™ H2O Moisture Filter

Item No	Description	Fittings
202230-B	New ZPure Moisture Filter	1/8" Brass
202230-SS		1/8" Stainless Steel
202231-B		1/4" Brass
202231-SS		1/4" Stainless Steel
202230R-B	Reconditioned ZPure Moisture Filter	1/8" Brass
202230R-SS		1/8" Stainless Steel
202231R-B		1/4" Brass
202231R-SS		1/4" Stainless Steel

Warnings

WARNING notices must be read carefully and understood. Improper use of this product can cause harm or death to personnel and damage to property!

WARNING

Remember to wear safety glasses.



Leak test all fittings when using hazardous or flammable gases.


Do not use above 14 bar (200psi).

Do not open filter cartridge.

Special precautions might be required when using hydrogen or other hazardous gasses. Consult local regulations and your company's safety procedures.

Limits:	Temperature	15°C to 35°C
	Pressure	0.75 to 14 bar
	Gas Supply	Inert gas or Hydrogen with < 50ppm O ₂ , H ₂ O and Hydrocarbons
	Flow Rate	< 2 L/min

Filter Setup and Use



1. Check the package contents for damage. Contact your supplier if any items are damaged or missing.
2. The ZPure filter is delivered in the active state, filled with helium. It is ready for use, but it is recommended that the trap be purged with the gas to be purified.
3. For the best connections, start with freshly-cut tubing in the gas line. 1/8" nuts should be turned 3/4 turn past finger tight. 1/4" nuts should be turned 1-1/4 turns past finger tight.
4. To connect in-line, first purge the gas line with the supply gas. While maintaining a low purge rate of 10 to 30 cc/min, remove the end plugs from both ends of the oxygen trap, then attach the inlet end of the trap to the gas line.

While the end plugs are removed from the trap during installation, a small amount of air will diffuse into the adsorbent, and the trap may warm slightly. As long as the trap is attached to the gas line within a few minutes, trap heating and the loss of oxygen-adsorbing capacity will be negligible.

5. Attach the outlet end of the trap to the instrument, then continue purging the entire system until a total of 1 to 2 liters of gas has been flushed through.
6. Leak check all connections.

Exchanging a Filter for a Reconditioned Filter

1. If you wish to return a used filter for reconditioning, please contact your supplier for return authorization and additional instructions.

Document Number 995105

Version 1

Printed in USA

Chromatography Research Supplies, Inc.
2601 Technology Drive
Louisville, KY 40299
USA

tel. (+1) 502-491-6300

www.chromres.com

Recycling



For recycling contact your CRS distributor