

ZPure™ Glass O₂ Filters with Quick Connect Fittings





Warnings

Warnings must be read carefully and understood. Improper use of this product can cause harm or death to personnel and damage to property!



Wear eye protection and use caution when working with pressurized systems.

Maximum pressure rating is noted in table below.

When using a high-pressure gas source overpressure protection must be provided.

Do not open filter, even after use.

The rapid oxidation of the active adsorbent in this filter can cause the filter bed to become extremely hot, resulting in danger of fire or burn injuries. Limited exposure of the filter to air during installation does not result in rapid reaction or dangerous heat generation.

Do not pass air or oxidizing gases in concentrations above 0.1% through this filter!


Do not expose the protective outer plastic tube to solvents.

Do not use without the protective outer plastic tube in place.


Special precautions might be required when using hydrogen or hazardous gases. Consult local regulations and your company's safety procedures.

Product Numbers

| Size | 1/8" Brass (Complete Kit) | 1/4" Brass (Complete Kit) | 1/8" SS (Complete Kit) | 1/4" SS (Complete Kit) | Replacement Cartridge (Brass) | Replacement Cartridge (SS) | Maximum Pressure |
|---------------------------|---------------------------------|---------------------------------|------------------------------|------------------------------|-------------------------------------|----------------------------------|----------------------|
| Glass 46 cc Filters | 202236-B-QC | 202237-B-QC | 202236-SS- QC | 202237-SS- QC | 202236RP-B-QC | 202236RP-SS- QC | 6.9 bar / 100 psi |



With Fittings



Without Fittings

Filter Setup and Use

Installation for Quick Connect Fittings

1. Check the package contents for damage. Contact your supplier if any items are damaged or missing.
2. The ZPure Filter is delivered in the active state, filled with helium. It is ready for use, but it is recommended that the trap be purged with the gas to be purified.
3. Install the Quick Connect Tube fittings ends to the gas lines. For the best connections, start with freshly-cut tubing in the gas line. 1/8" nuts should be turned 3/4 turn past finger tight. 1/4" nuts should be turned 1-1/4 turns past finger tight.
4. Attach the outlet end of the filter by screwing the Quick Connect fitting nut onto the adapter. This will open both check valves on the outlet side.
5. Attach the inlet end of the filter by screwing the Quick Connect fitting nut onto the adapter. This will open both check valves on the inlet side.
6. Purge the entire system until a total of 1 to 2 liters of gas has been flushed through.
7. Leak check all connections.

Chromatography Research Supplies, Inc.
2601 Technology Drive
Louisville, KY 40299
USA

tel. (+1) 502-491-6300

www.ChromRes.com



Recycling

Please contact CRS for information regarding recycling and reconditioning gas purifier products.

Printed in USA
28-Sep-2017
995207